Melt Flow Indexer

Model: M0004

The Melt Flow Indexer is a dead-weight extrusion plastometer. It consists of a thermostatically controlled melting chamber (the barrel) in which the polymer under test is heated and from which it is extruded through a standard die under standard conditions of load, which is, of course, made up of the combined weights of the extrusion piston and the loose weight both of which are carefully calibrated to well within the most stringent limits.

1DV instruments

1DM

Applications:

Plastics

Standards:

- BS2782
- ASTM D1238: Procedure A
- ISO 1133

Features:

Barrel: Precision ground and honed.
 Manufactured from high grade tool steel for longer life.

50.8 mm outside diameter 9.55 mm inside diameter 162 mm long

 Temperature: The temperature of the barrel is controlled by the Precision Digital Temperature Controller.

Temperature Range: 100° - 300°C, ± 0.2°C. 100° - 400°C, ± 0.2°C.

• Piston:

Weight - 100 gm.
• Piston Weights:
4.9 Kg weight (1)
2.06 Kg weight (1)

Diameter: 9.47 mm

- Sample Cut Off Knife (1)
- Die Remover (1)
- Die Cleaner (1)
- Cleaning Tool (1)
- Filler Tool (1)
- Level (1)

Optional:

Die: Tungsten Carbide:
8 mm long
9.5504 mm overall diameter
2.0955 mm internal diameter
The alternative Standard die in BS 2782 method 1050 with a bore diameter of
1.181 mm is also available on request.

• Piston:

Diameter: 9.47 mm Weight - 100 gm

- Calibration Certificate
- Spare Die

Working Environment:

• Operating Temperature: -10°C to 55°C

(with no condensation or icing)

• Operating Humidity: 25% to 65%

• Storage Temperature: -25°C to 65°C

(with no condensation or icing)

Melt Flow Indexer

Model: M0004

Connections:

• Electrical: 220/240 VAC @ 50 HZ or

110 VAC @ 60 HZ

(please specify when ordering)



Dimensions:

H: 480mmW: 430mmD: 270mmWeight: 27kg

Procedural Conditions:

Material	Condition	
Acetals (copolymerand homoploymer)	190/2.16	190/1.05
Acrylics	230/1.2	230/3.8
Acrylonitrile-butadiene-styrene	200/5.0	230/3.8
	220/10	
Acrylonitrile/butadiene/styrene/polycarbonate	230/3.8	250/1.2
blends	265/3.8	265/5.0
Cellulose esters	190/0.325	190/2.16
	190/21.60	210/2.16
Ethylene-chlorotrifluoroethylene copolymer	271.5/2.16	
Ethylene-tetrafluoroethylene copolymer	297/5.0	
Nylon	275/0.325	235/1.0
	235/2.16	235/5.0
	275/5.0	
Perfluoro (ethylene-propylene) copolymer	372/2.16	
Perfluoroalkoxyalkane	372/5.0	
Polycaprolactone	125/2.16	80/2.16
Polychlorotrifluorethylene	265/12.5	
Polyethylene	125/0.325	125/2.16
	250/1.2	
	190/0.325	190/2.16
	190/21.60	190/10
	310/12.5	
Polycarbonate	300/1.2	
Polymonochlorotrifluoroethylene	265/21.6	
	265/31.6	
Polypropylene	230/2.16	
Polystyrene	200/5.0	230/1.2
	230/3.8	190/5.0
Polyterephthalate	250/2.16	210/2.16
	285/2.16	
Poly (vinyl acetal)	150/21.6	
Poly (vinylidene fluoride)	230/21.6	
	230/5.0	
Poly (phenylene sulfide)	315/5.0	
Styrene acrylonitrile	220/10	230/10
	230/3.8	
Styrenic Thermoplastic Elastomer	190/2.16	200/5.0
Thermoplastic Elastomer-Ether-Ester	190/2.16	220/2.16
	230/2.16	240/2.16
		250/2.16
Thermoplasyic Elastomers (TEO)	230/2.16	
Vinylidene fluoride copolymers	230/21.6	
	230/5.0	

These conditions have been found satisfactory for the materials listed

Web: www.idminstruments.com.au